



**KERN & Sohn GmbH**

Ziegelei 1

D-72336 Balingen

E-Mail: [info@kern-sohn.com](mailto:info@kern-sohn.com)

Tel: +49-[0]7433- 9933-0

Fax: +49-[0]7433-9933-149

Internet: [www.kern-sohn.com](http://www.kern-sohn.com)

# Installation Information

## Rechargeable battery

### **KERN WTB-A01N**

Version 1.0

11/2012

GB

WTB-N-BA-e-1220



# KERN WTB-A01N

Version 1.0 11/2012


## Installation Information Rechargeable battery

---


---

**The rechargeable battery is charged via the power supply of the balance.**  
The operating time of the rechargeable battery with background illumination is 30h, without that it is 50 hrs. Charging time until complete recharging 12h.


### Insert rechargeable battery:



- ⇒ Remove any adherent moisture from the balance
- ⇒ Rotate both levers (1) 90 degrees to the left



⇒ Remove battery compartment cover and the foam pieces



⇒ Connecting the rechargeable battery



Watch the colours:

Red on red !


Black on black!




⇒ Insert the rechargeable battery in a manner that it cannot slip (fasten with a rubber foam mat)



Ensure that the cables are not squeezed.



- ⇒ Close the rechargeable battery cover
- ⇒ Rotate both levers (1) 90 degrees to the right

If in the display the triangle ▼ appears above the capacity display , the rechargeable battery is nearly exhausted. Connect power pack, the rechargeable battery is loaded.



Avoid excess pressure on the balance in order to avoid damaging it, particularly as it is supported by the weighing plate.